

SOLID HYDROGEN RAMAN SHIFTER FOR MID-INFRARED RANGE (4.4 – 8 μm)

KIRILL E. KUYANOV, HIROMICHI HOSHINA, and ANDREY F. VILESOV, *Department of Chemistry, University of Southern California, Los Angeles, CA 90089*; TAKAMASA MOMOSE, *Division of Chemistry, Graduate School of Science, Kyoto University, Kyoto, 605-8502, Japan*.

We have developed a pulsed continuously tunable laboratory laser source for the mid-infrared spectral range of 4.4 – 8 μm , having a spectral line width of 0.4 cm^{-1} . The device is based on the backward stimulated Raman scattering (SRS) of the focused near-infrared (1.56 – 1.85 μm) laser beam in solid *para*-hydrogen at $T = 4\text{ K}$. The output energies range from 1.7 mJ at 4.4 μm to 120 μJ at 8 μm , which corresponds to SRS quantum efficiencies ranging from 0.53 to 0.17, respectively.